

# **User regulations**

## **for users of university institutions**

### **Core Facility for Cell Sorting and Cell Analysis**

Version of 29<sup>th</sup> June 2022

#### **1 Organization**

##### **1.1 Management**

Prof. Dr. Brigitte Müller-Hilke

Core Facility for Cell Sorting and Cell Analysis

Schillingallee 70, 18057 Rostock

Email: [brigitte.mueller-hilke@med.uni-rostock.de](mailto:brigitte.mueller-hilke@med.uni-rostock.de)

Tel .: +49 381 494 5883

##### **1.2 Technical management**

Wendy Bergmann, M.Sc.

Core Facility for Cell Sorting and Cell Analysis

Schillingallee 70, 18057 Rostock

Email: [wendy.bergmann@med.uni-rostock.de](mailto:wendy.bergmann@med.uni-rostock.de)

Tel .: +49 381 494 5876

##### **1.3 Research Associate**

Marcel Kordt, M.Sc.

Core Facility for Cell Sorting and Cell Analysis

Schillingallee 70, 18057 Rostock

Email: [marcel.kordt@med.uni-rostock.de](mailto:marcel.kordt@med.uni-rostock.de)

Tel .: +49 381 494 5877

## 1.4 Access

Information about the Core Facility for Cell Sorting and Cell Analysis can be found on the internet at <https://zsa.med.uni-rostock.de>.

The offer is available to the following user groups:

- University medical staff
- Employees of the University of Rostock
- External users (separate user regulations, see “User regulations for users of non-university institutions”)

At full capacity, the staff members decide whether to accept new projects, whereby groups from the Rostock University Medical Center will be given priority.

The employees of the Core Facility are usually available between 9 a.m. and 5 p.m. After having obtained complete instruction on the devices, the users are free to use them in the specified time and after consultation with the responsible laboratory managers. These are Prof. Müller-Hilke for the complex 2.67-2.70 and PD Dr. Kirsten Peters for the ZEMFO 1.030A. Access to the above-mentioned rooms is granted by applying to the laboratory management in the form of locking authorization.

## 2 Equipment

The Core Facility for Cell Sorting and Cell Analysis has three flow cytometers for analysis; the Cytek™ Aurora, the BD FACSCalibur™ and the BD FACSVerse™, three different cell sorting systems; BD FACSAria™ IIIu, the autoMACS® Pro Separator, as well as magnets for manual MACS separations and the Luminex100 / 200 for the multiplex bead-based analysis of soluble proteins. A USB dongle with a transportable license for the FlowJo™ software can be borrowed for data evaluation. The exact description of the devices, their configuration and applicability can be found on the homepage <https://zsa.med.uni-rostock.de/ausstattung>.

## 3 Range of services

### 3.1 Services offered by the Core Facility

- Assistance in planning of flow cytometric experiments
- Individual instruction on the devices and the acquisition software
- Advice on possibilities of cell sorting and their implementation on the BD FACSAria™ IIIu
- 
- Establishment and measurement of marker panels (up to 24 colors)
- Advice and implementation of magnetic-based cell separation
- Support in the evaluation of flow cytometric data including advice on the subsequent statistics
- Regular care and maintenance of the devices, performance check and data backup

### 3.2 Obligation of the user

- Recognition of the user regulations by the project manager and the user him/herself
- Booking of devices via the online booking portal *bookkit*
- Responsible use of the devices and provided workplaces

- Transport and anonymization of samples to the premises of the Core Facility
- Information about the success of experiments, especially after sorting

## **4 Registration**

Individuals interested in using the Core Facility contact its employees to make an appointment for an initial meeting. This initial meeting serves to discuss the project at hand, the suitability of devices, sample preparation and protocol.

Up to three appointments (approx. 1.5 hours each) - depending on the complexity of the device and the individual prior knowledge - are agreed upon during which the user is familiarized with the device and the corresponding software. The user carries out the device settings for the specific sample with the help of employees.

Interested users acknowledge our user regulations, which can be downloaded from the Core Facility homepage. The user regulations are printed out and signed by the project manager and the user, then returned to the employees.

The user receives an email from the online booking portal bookkit. After successful registration and completion of the training units, the user is authorized by the employees to book the devices and consumables via bookkit

This procedure is summarized in a welcome email that the interested user receives after the initial meeting.

In addition, the user obtains an email with some links and tips on the theory of flow cytometry.

## **5 Booking of devices**

After successfully completing the instruction, the user is authorized by the staff to book the devices via bookkit. The BD FACSAria™ IIIu is an exception, its use is only possible after direct consultation with the employees of the Core Facility.

If any device is defective, this will be shown in the calendar, as a notice on the homepage and a notification will be sent to the user via email. Users affected by a defective device report immediately to the employees of the Core Facility, so that the measurements can be transferred to an alternative device.

Booking by or for colleagues is not allowed. However, an unregistered user can work on the devices under the supervision of a registered, experienced user. This must be expressly approved by the staff of the Core Facility.

## **6 Cancellation policies**

Booked sessions can be canceled up to 24 hours before the planned use of the device. The user is asked to inform subsequent users about this case so that they can bring forward their measurements if necessary. Cancellations shorter than 24 hours will be charged at the full hourly rate.

If the last booking on that day is canceled at very short notice, the user must ensure that the device is switched off correctly.

## **7 Responsibility of the user**

The user agrees to the user regulations via his/her own signature plus the signature of his/her supervisor and submits the signed documents to the employees of the Core Facility. The user may only use the corresponding devices independently after successful instruction, after returning the signed user regulations and after having received safety instructions. Each user is responsible for cleaning the devices according to the instructions, for setting them up for the next measurement, for documenting the session in the logbook and for cleaning the workplace (tidy up and disinfect).

## **8 Responsibility of the Core Facility**

The Core Facility employees service and clean the devices regularly. A so-called performance check is performed regularly and ensures reliable measurements. Every activity by the employees is documented in the folder for users to view. The employees maintain discretion about planned and ongoing projects towards third parties. Exceptions must be agreed individually.

## **9 Biosafety**

The user's supervisor is responsible for the annual instructions on genetic engineering, infection protection, occupational safety, etc. and guarantees the user's knowledge on the respective safety classifications, transport and handling of biological samples.

As part of the briefings, a short safety instruction is given by the employees of the Core Facility. This is documented with a signature (see appendix to this user regulation) and is repeated annually. The laboratories in which the devices are located are S2 approved. Appropriate protective measures, such as wearing laboratory protective clothing (bring your own protective clothing or take it out of the central pool) and hygienic working methods (disinfection of workplace and hands) are compulsory.

The users must document their measurements in the logbook, which is available on each of the devices. Especially measurements of S2-level samples (e.g. human samples, virally transduced samples) must be reported.

Higher safety measures must be taken when using the BD FACSAria™ IIIu, as the device produces aerosols due to its operating modes. Therefore it is compulsory that the employees of the Core Facility are informed in advance of the safety aspects of the sample to be analyzed or sorted. If necessary, appropriate measures for prevention and decontamination of the device are discussed and initiated.

## **10 Data collection evaluation and storage**

The Core Facility is responsible for the correctness of the data collection and evaluation. The user is obliged to export his or her data and save it on an external storage medium. The collected data belongs entirely to the user. The standard formats fcs, xls, doc, odt, ods, csv, png and svg can be used for data exchange.

The employees archive the raw data every six months in the form of a backup, which saves the data on external hard drives of the appropriate devices and also on external, transportable hard drives as part of the regular cleaning / maintenance of the devices.

## 11 Acknowledgments in publications

The users of the core facility agree to adequately recognize the use of the core facility in accordance with the rules of good scientific practice. This means that the use of the devices and the services provided by the Core Facility staff must be mentioned in every form of publication. This also includes theses, conference contributions and third-party funding applications.

Employees of the Core Facility will be taken into account when publishing data that was established and collected in the Core Facility with the cooperation of the responsible persons of the Core Facility. Co-authorship is granted depending on the scope of services provided by the respective employees. Of course, this requires participation in the writing of the corresponding manuscript.

For the Core Facility, the above-mentioned references are important proof of the necessity for their existence and serve the further development within the framework of evaluation processes and third-party funding applications.

## 12 Costs

Consumables such as sheath, cleaning fluids and the beads for the respective quality controls are billed by the users as part of an hourly rate. Small test volumes of reagents and AKs can be purchased from the Core Facility.

A price list for the use of the device and the consumables offered follows:

Device name	Hourly price or separation price (gross) *
BD FACSAria™ IIIu	once 50,00 € plus 3,00 € per hour
BD FACSVerser™	4,50 €
BD FACSCalibur™	2,50 €
Cytek®Aurora	3,00 €
autoMACS® Pro Separator	11,00 €*
Luminex®100/200™	once 35,00 € plus 4,00 € per plate

Reagent	Producer	Gross price per test/ piece
7-AAD Viability Staining Solution	BioLegend # 420403	0,14 €
DAPI (1:3600)	BioLegend # 422801	0,03 €
Propidiumiodid (10 µL)	eBiosciences # BMS500PI	0,67 €
ZombieNIR	BioLegend # 423105	0,80 €
ZombieRed	BioLegend #423109	0,80 €
LIVE/DEAD Fixable Aqua Viability Dead Stain Kit	Invitrogen # L34957	2,24 €
Fixable Viability Dye eFluor506	eBiosciences #65-0866-14	1,45 €
Fixable Viability Dye eFluor660	eBiosciences # 65-0864-14	1,48 €
Fixable Viability Stain 700	BD Horizon # 564997	0,89 €
UltraComp eBeads	Invitrogen # 01-2222-42	4,19 €
UltraComp eBeads Plus	Invitrogen #01-3333-42	4,31 €
MACS Comp Bead anti-REA	Miltenyi Biotec #130-104-693	1,25 €
5 ml Tube PP with lid (sterile)	Falcon™ 352063	0,30 €
5 ml Tube with cell sieve (sterile)	StemCell # 100-0087	3,57 € or 89,25 € / 25 pieces
5 mL Tube PS (non-sterile)	Sarstedt 551579	0,09 €

Antibodies (human, murin)	variable	Producer's list price
Lactadherin:FITC	Haematologic Technologies # BLAC-FITC	4,48 €
Apotracker Green	BioLegend # 427402	2,20 €
CFSE Cell Division Tracker Kit	BioLegend # 423801	0,02 € / 5 µM

Fees for the use of the device and additionally requested consumables are billed quarterly and sent to the project managers with the request to transfer the costs.

user name: \_\_\_\_\_

user group: \_\_\_\_\_

## Recognition by signature

Rostock, 29<sup>th</sup> June 2022

Date



Signature Prof. Dr. Brigitte Müller-Hilke, Head of the core facility

Rostock, \_\_\_\_\_

Date

Signature user

Rostock, \_\_\_\_\_

Date

Signature project manager

**Appendix**

**Documentation of safety instructions for the user**

Date	User	Signature	Instructor